

# Notice of Allowability

Application No.

10/828,540

Examiner

L. G. Lauchman

Applicant(s)

FUJIMOTO ET AL.

Art Unit

2877

## -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 11/01/2007.
2. ☒ The allowed claim(s) is/are 1,2,5-10,14 and 15.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☒ All    b) ☐ Some\*    c) ☐ None    of the:
    1. ☒ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  5. ☐ CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

### Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO/SB/08),  
Paper No./Mail Date 11/01/2007
4. ☐ Examiner's Comment Regarding Requirement for Deposit  
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),  
Paper No./Mail Date \_\_\_\_\_
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_

**DETAILED ACTION**

***Allowable Subject Matter***

Claims 1, 2, 5-10, 14, and 15 are allowed.

The following is an examiner's statement of reasons for allowance:

As to Claim 1, the prior art of record taken alone or in combination, fails to disclose or render obvious the first light beams are wider along a direction perpendicular to the scanning direction than along a direction parallel to the scanning direction, and the first-side light detecting device has a light detecting area that is wider along the direction perpendicular to the scanning direction than along the direction parallel to the scanning direction, and the second light beams are wider along a direction perpendicular to the scanning direction than along a direction parallel to the scanning direction, and the second-side light detecting device has a light detecting area that is wider along the direction perpendicular to the scanning direction than along the direction parallel to the scanning direction, in combination with the rest of the limitations of the claim.

As to Claim 5, the prior art of record taken alone or in combination, fails to disclose or render obvious the first light beams are wider along a direction perpendicular to the scanning direction than along a direction parallel to the scanning direction, and the first-side light detecting device has a light detecting area that is wider along the direction perpendicular to the scanning direction than along the direction parallel to the scanning direction, and the second light beams are wider along a direction perpendicular to the scanning direction than along a direction parallel to the scanning direction, and the second-side light detecting device has a light detecting area that is wider along the direction perpendicular to the scanning direction than along the

direction parallel to the scanning direction, in combination with the rest of the limitations of the claim.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to L. G. Lauchman whose telephone number is (571) 272-2418. The examiner's normal work schedule is 8:00am to 4:30pm (EST), Monday through Friday. If attempts to reach examiner by the telephone are unsuccessful, the examiner's supervisor Gregory J. Toatley, Jr. can be reached on (571) 272-2059, ext. 77.

The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Application/Control Number:  
10/828,540  
Art Unit: 2877

Page 4

Any inquiry of a general nature or relating to the status of this application should be directed to the TC receptionist whose telephone number is (571) 272-1562.

A handwritten signature in black ink, appearing to read 'L. G. Lauchman', with a large, stylized initial 'L' and a long horizontal stroke extending to the right.

L. G. Lauchman  
Primary Examiner  
Art Unit 2877

11/13/2007